

**Protocol of the intercomparison of the Fourier transform  
spectroradiometer PFS of the PTB with the travelling reference  
spectroradiometer QASUME from PMOD/WRC at PTB Berlin,  
Germany, on March 31 to April 4, 2014**

Report prepared by Gregor Hülsen and Peter Meindl

Operator of QASUME: Gregor Hülsen

Operators of PFS: Peter Meindl and Martin Wähmer

The purpose of the visit was the comparison of global solar irradiance measurements between the spectroradiometer PFS operated by PTB Berlin and the travel reference spectroradiometer QASUME. The measurement site is located at Berlin; Latitude 52.52 N, Longitude 13.32 E and altitude 59 m.a.s.l. The horizon of the measurement site is free down to at least 85° solar zenith angle (SZA). Measurements between 4:45 UT and 17:30 UT have been analysed.

QASUME was installed on the roof of the PTB building "Siemensbau" in the afternoon of March 31, 2014. The spectroradiometer was installed next to the PFS spectroradiometer with the entrance optic of QASUME within 0.5 m to the other instrument. The spectroradiometer in use at PTB is a Fourier transform spectroradiometer (FTS) prototype. The input optics is a J1002 from CMS Schreder – identical to the QASUME optic. The intercomparison between QASUME and the PFS spectroradiometer lasted four days, from morning of April 1<sup>st</sup> to the afternoon of April 4<sup>th</sup>.

QASUME was calibrated several times during the intercomparison period using a portable calibration system. Three lamps (T61253, T68522 and T685240) were used to obtain an absolute spectral irradiance calibration traceable to the primary reference held at PMOD/WRC, which is traceable to PTB. The daily mean responsivity of the instrument based on these calibrations varied by less than  $\pm 1\%$  during the intercomparison period. The internal temperature of QASUME was  $(24.3 \pm 0.1)$  °C and the diffuser head was heated to a temperature 31.7 °C. Due to a malfunction of the diffusor head temperature sensor the head temperature could not be monitored during the afternoon of April 1<sup>st</sup>.

The wavelength shifts relative to an extraterrestrial spectrum as retrieved from the SHICRivm analysis were between  $\pm 50$  pm in the spectral range 290 to 400 nm.

The Fourier transform spectroradiometer PFS was also calibrated several times during the intercomparison period using the QASUME portable calibration system. Further calibrations have been performed in advance against a high-temperature black body radiator and against a secondary irradiance standard of PTB which is traceable to the national irradiance standard of PTB. The diffuser head temperature was heated to 28,0 °C. However, this temperature was not monitored. The temperature of the FTS housing was stabilized to 25 °C.

### Protocol:

The measurement protocol was to measure solar irradiance spectra from 290 nm to 400 nm. QASUME measured one spectrum every 15 minutes, in steps of 0.25 nm and 1.5 seconds between each wavelength increment. The FTS recorded continuously interferograms where each interferogram was measured in about 0.6 sec. In order to limit the amount of data, 32 interferograms recorded in 19 seconds have been averaged and saved consecutively. A further averaging of the recorded interferograms have been performed during the postprocessing of the data depending on the atmospheric conditions. The spectral resolution of the FTS was set to  $10\text{ cm}^{-1}$ . A Blackman-Harris-4-Term apodization was applied to the measured interferograms to reduce the signal leakage into side lobes. The resolution after applying this apodization is about  $20\text{ cm}^{-1}$ . A Hamamatsu photosensor module H10723-210 in combination with different spectral filters has been applied during the measurements.



DOY	Date	DAY	Weather	Comment (times are in UT)
90	31. Mrz	Monday	Mix of Sun & Clouds	Installed at 12 UT First Scans at 13:45 16 UT: FTS calibration (T61253) FTS Filter: UG5
91	01. Apr	Tuesday	Mix of Sun & Clouds, Haze	9:13 calibration (T68522) 16:30: FTS calibration (T61253) FTS Filter: UG5 16:45: FTS calibration (T61253) FTS Filter: UG11 9:30-15:30: No head temp. Control
92	02. Apr	Wednesday	Mix of Sun & Clouds, Haze	16:44 calibration (T685240) 17:30 FTS Filter UG5
93	03. Apr	Thursday	Clear sky with few Cirrus Mix of S&C in the afternoon	6:59 calibration (T685240) 7:15 calibration (T61253)
94	04. Apr	Friday	Mix of sun & clouds	13:45+14:03 calibration T61253+T68522  End of Campaign: 15:00




### Results:





In total 159 spectra from QASUME and 8005 PFS interferograms are available from the measurement period. Measurements between 4:45 and 17:30 UT have been analysed (SZA smaller than  $90^\circ$ ).



**Remarks:**

1. The interferograms of the FTS were processed as explained above.
2. The resulting spectra are further processed using the “matshic.m” algorithm. Finally the spectra of the FTS recorded during QASUME scan were extracted using “griddata.m” to enable a comparison of the two datasets.
3. Standard Lamp calibrations:
  - a. The PFS spectroradiometer was calibrated by the PTB using a 250 W transfer standard KSL003 and a HTBB3200pg (High Temperature Blackbody) in the laboratory of the PTB before the intercomparison.
  - b. Monday evening, 1<sup>st</sup> April, the PFS system was calibrated using the T61253 250 W transfer standard from PMOD/WRC in the FTS configuration 1 (UG5 filter). This measurement was repeated 24h later (2<sup>nd</sup> April evening). The calibration was also performed in the FTS configuration 2 (UG11-UV filter).
  - c. The difference of the PTB and the PMOD-Lamp calibrations are around 3%
  - d. The final dataset of PTB is based on the T61253 calibration.
4. Two spectral filter configurations inside the FTS were tested during the campaign:
  - a. 1<sup>st</sup>, 3<sup>rd</sup> and 4<sup>th</sup> April: UG5 filter
  - b. 2<sup>nd</sup> April: UG11 filter
5. The ratios between PFS and QASUME have on average an offset of -5 %.
6. The diurnal variation of the PFS to QASUME ratio is 8 %. The ratios PFS to QASUME decrease constantly from the morning to the afternoon and rise again until the evening. The reason can to some degree be addressed to the different input optic of the two devices (Azimuth Error, Temperature / Humidity Sensitivity of the Teflon Sensor, Cosine Error).
7. For all solar scans the wavelength shifts of the PFS are 0 pm.
8. The original dataset has been revised:
  - a. Off-axis-light correction
  - b. Vacuum-Air Wavelength conversion
  - c. New PFS responsivity. Adjusted for artefacts from the HeNe-laser line at 316.5 nm. Vacuum-Air Wavelength conversion.
9. Differences using the UG5 and UG11 filter:
  - a. UG5 Dataset: A prominent peak around 316.5 nm for low irradiance levels (SZA<60 deg) – see figures. This is an artefact caused by the internal HeNe laser of the PFS.
  - b. UG11 Dataset (only DOY 93): The peak at 316.5 nm is not visible. However, the ratios become noisy for wavelengths >390 nm.









Global irradiance ratios PFS/QASUME at PTB-matshic:01-Apr-2014(091)

