# Protocol of the intercomparison at Grossenzersdorf, BOKU, Austria from April 29 to May 2, 2024 with the travelling reference spectroradiometer QASUME from PMOD/WRC

Report prepared by Julian Gröbner

Operator: Julian Gröbner

Local operators: Daniel Rauter and Stana Simic

The purpose of the visit was the comparison of global solar irradiance measurements between the spectroradiometer operated by BOKU and the travel reference spectroradiometer QASUME. The measurement site is located at Grossenzersdorf; Latitude 48.200° N, Longitude 16.559° E and altitude 159 m.a.s.l.

The horizon of the measurement site is free down to at least 85° solar zenith angle (SZA). Measurements between 4:30 UT and 17:30 UT were analysed.

QASUME was installed on the roof of the measurement container of BOKU in the morning of April 29, 2024. The spectroradiometer was installed next to the entrance optic of BOKU "GSA" with the entrance optic of QASUME within less than 2 m of "GSA". The spectroradiometer "GSA" is a double monochromator of type IDR-150 of Bentham. The intercomparison between QASUME and GSA lasted 2 1/2 days, from the afternoon of Monday April 29 to the morning of Thursday May 2.

QASUME was calibrated several times during the intercomparison period using a portable calibration system. Three lamps (T61251, T68523, and T157825) were used to obtain an absolute spectral irradiance calibration traceable to the primary reference held at PMOD/WRC, which is traceable to PTB. The daily mean responsivity of the instrument based on these calibrations varied by less than 1 % during the intercomparison period. The internal temperature of QASUME was  $27.1\pm0.3$  °C and the diffuser head was heated to a temperature of  $28.9\pm0.6$  °C.

The wavelength shifts relative to an extraterrestrial spectrum as retrieved from the matSHIC analysis were between ±50 pm in the spectral range 300 to 500 nm.

### **Protocol:**

The measurement protocol was to measure one solar irradiance spectrum every 30 minutes from 290 nm to 500 nm, every 0.5 nm, and 3.0 seconds between each wavelength increment.

DOY	Date	DAY	Weather	Comment (times are in UT)
120	29-Apr	Monday	Clear Sky with cirrus, sahara dust, very diffuse	Installed at 8:00 10:00 start UV measurements
121	30-Apr	Tuesday	Clear sky, small Cu developing in the afternoon	Calibrated at 12:15 UT  11:45 Calibration
122	01-May	Wednesday	Mix of sun and clouds, windy	Calibration at 13:45
123	02-May	Thursday	Overcast, rain	7:30 UT QASUME OFF

#### **Results:**

In total 64 synchronised simultaneous spectra from QASUME and GSA are available from the measurement period. Measurements between 4:30 UT and 17:30 UT have been analysed (SZA smaller than 85°).

#### **Conclusions:**

- 1. The spectral irradiances measured by GSA are on average 1.0% higher than those of QASUME for wavelengths shorter than 400 nm, and 2.7% higher at wavelengths longer than 400 nm.
- 2. The temporal variation of the spectra between GSA and QASUME was very stable, to within 2% for wavelengths shorter than 400 nm. At longer wavelengths, a diurnal variation of up to 5% could be seen in the SZA range between approximately 50° and 75°, which could be due to a non accounted cosine error.
- 3. The wavelength shifts of GSA relative to the high spectral resolution solar spectrum TSIS-1 HSRS are between ±50 pm in the spectral range 300 nm to 500 nm.
- 4. The signal to noise ratio at wavelengths shorter than about 305 nm decreases with increasing SZA.

## Comparison to previous QASUME site visits

The previous QASUME visit was in 2004, see report at <a href="https://www.pmodwrc.ch/wcc\_uv/qasume\_audit/reports/2004\_05\_austria\_grossenzersdorf\_ATW1.pdf">https://www.pmodwrc.ch/wcc\_uv/qasume\_audit/reports/2004\_05\_austria\_grossenzersdorf\_ATW1.pdf</a>

### **Comments from the operator:**

The BOKU GSA instrument was calibrated before the QASUME intercomparison using the primary standard and twice during the campaign using the secondary standard.

BOKU-Met uses a NIST 1000 W FEL primary standard for calibrations and a 1000 W secondary standard for more regular (monthly) calibrations.

The secondary standard lamp sits inside a self-built housing which fits on top of the entrance optics, so distance, temperature stabilisation and shielding from outside light is provided.

Lamps are powered with a high precision power supply and temperature stabilisation is provided by holding a constant of voltage (+/- 0.05 mV DC) on a shunt resistor.

After a defect of the HV power supply of the instrument, it was sent to GB for repairs and upgrade (from DM-150 to IDR-150). Since 2022 it is again in continuous operation.

The instrument is located inside a container which is temperature stabilised to approximately  $\pm 2.5$  °C.

The envirobox of the instrument is cooled with a high-precision cooling system (heat exchanger), custom built for BOKU-Met.

Thus the PMT temperature is stabilised to  $20 \pm 0.2$  °C.





















